

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110220\
 Data File : VN064420.D
 Acq On : 2 Nov 2020 17:37
 Operator : JC/MD
 Sample : L4576-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PW-2

Quant Time: Nov 02 17:58:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N110220W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Nov 02 15:45:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	53904	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	290766	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	260000	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	7.98	65	139403	31.68	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	105.60%
60) 4-Bromofluorobenzene	12.37	95	112395	25.58	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	85.27%
63) Toluene-d8	10.06	98	349394	30.14	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.47%

Target Compounds

					Ovalue
25) Chloroform	7.33	83	308442	46.647 ug/l	97
41) Bromodichloromethane	9.37	83	49275	9.465 ug/l	97
53) Dibromochloromethane	10.87	129	7870	1.936 ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110220\
Data File : VN064420.D
Acq On : 2 Nov 2020 17:37
Operator : JC/MD
Sample : L4576-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PW-2

Quant Time: Nov 02 17:58:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N110220W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Nov 02 15:45:48 2020
Response via : Initial Calibration

