


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN110220\
 Data File : VN064410.D
 Acq On : 2 Nov 2020 11:36
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N110220W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Tue Nov 03 02:03:38 2020



AutoFind: Scans 3306, 3307, 3308; Background Corrected with Scan 3294

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.0	12764	PASS
75	95	30	60	53.9	32741	PASS
95	95	100	100	100.0	60706	PASS
96	95	5	9	6.7	4092	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	79.7	48413	PASS
175	174	5	9	8.4	4088	PASS
176	174	95	101	96.6	46760	PASS
177	176	5	9	5.8	2704	PASS