

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110221\
 Data File : VN069342.D
 Acq On : 2 Nov 2021 12:33
 Operator : JC/MD
 Sample : M4438-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FDR-20211101-WC-LM

Quant Time: Nov 03 00:47:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	296613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	531184	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	545408	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	226920	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	204561	47.775	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.560%
35) Dibromofluoromethane	8.024	113	154644	46.703	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.400%
50) Toluene-d8	10.443	98	669638	50.580	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.160%
62) 4-Bromofluorobenzene	12.731	95	264037	52.851	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.700%
Target Compounds						
					Qvalue	
16) Acetone	4.301	43	28478	15.611	ug/l	96
17) Carbon Disulfide	4.527	76	5972	0.841	ug/l	99
25) 2-Butanone	7.348	43	8223	4.556	ug/l	99
40) Benzene	8.469	78	5933	0.431	ug/l	97
56) Ethyl methacrylate	10.762	69	6542	2.633	ug/l	96
83) tert-Butylbenzene	13.326	119	71569	5.180	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110221\
Data File : VN069342.D
Acq On : 2 Nov 2021 12:33
Operator : JC/MD
Sample : M4438-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FDR-20211101-WC-LM

Quant Time: Nov 03 00:47:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

