

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110221\
 Data File : VN069344.D
 Acq On : 2 Nov 2021 13:25
 Operator : JC/MD
 Sample : M4438-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 D20211101

Quant Time: Nov 03 00:47:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 28 16:09:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	279258	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	505479	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	516417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	214425	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	194716	48.302	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.600%		
35) Dibromofluoromethane	8.024	113	148193	47.031	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.060%		
50) Toluene-d8	10.443	98	634336	50.350	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.700%		
62) 4-Bromofluorobenzene	12.733	95	248280	52.224	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	104.440%		
Target Compounds					Qvalue	
16) Acetone	4.301	43	20717	12.062	ug/l	96
17) Carbon Disulfide	4.537	76	9183	1.373	ug/l	98
25) 2-Butanone	7.351	43	7023	4.285	ug/l	98
40) Benzene	8.461	78	6356	0.485	ug/l	95
56) Ethyl methacrylate	10.765	69	6065	2.609	ug/l	89
83) tert-Butylbenzene	13.326	119	76064	5.826	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110221\
Data File : VN069344.D
Acq On : 2 Nov 2021 13:25
Operator : JC/MD
Sample : M4438-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
D20211101

Quant Time: Nov 03 00:47:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102821W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 28 16:09:15 2021
Response via : Initial Calibration

