

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
 Data File : VN075110.D
 Acq On : 02 Nov 2022 12:16
 Operator : JC\MD
 Sample : VN1102WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1102WBL01

Quant Time: Nov 03 04:14:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	92430	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	177311	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	152636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	56978	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	76323	61.285	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 122.560%	
35) Dibromofluoromethane	8.171	113	57230	54.981	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.960%	
50) Toluene-d8	10.571	98	213961	52.981	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.960%	
62) 4-Bromofluorobenzene	12.853	95	64163	45.229	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.460%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
Data File : VN075110.D
Acq On : 02 Nov 2022 12:16
Operator : JC\MD
Sample : VN1102WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1102WBL01

Quant Time: Nov 03 04:14:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

