

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
 Data File : VN075113.D
 Acq On : 02 Nov 2022 13:39
 Operator : JC\MD
 Sample : PB148713TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148713TB

Quant Time: Nov 03 04:15:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	98527	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	184652	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	174212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	66706	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	76923	57.944	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 115.880%			
35) Dibromofluoromethane	8.177	113	59816	55.181	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 110.360%			
50) Toluene-d8	10.570	98	230922	54.908	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 109.820%			
62) 4-Bromofluorobenzene	12.853	95	75682	51.228	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 102.460%			
Target Compounds						Qvalue
16) Acetone	4.441	43	10939	19.461	ug/l	90
20) Methylene Chloride	5.288	84	6300	5.027	ug/l	94
37) Ethyl Acetate	7.577	43	14727	8.002	ug/l	96
43) Isopropyl Acetate	8.694	43	66582	23.323	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
Data File : VN075113.D
Acq On : 02 Nov 2022 13:39
Operator : JC\MD
Sample : PB148713TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148713TB

Quant Time: Nov 03 04:15:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

