

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
 Data File : VN075115.D
 Acq On : 02 Nov 2022 14:27
 Operator : JC\MD
 Sample : N5380-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-D

Quant Time: Nov 03 04:16:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	98529	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	192533	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	180762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	69336	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	77973	58.734	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.460%
35) Dibromofluoromethane	8.177	113	59685	52.806	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.620%
50) Toluene-d8	10.571	98	240952	54.948	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.900%
62) 4-Bromofluorobenzene	12.853	95	75613	49.086	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.180%
Target Compounds						
						Qvalue
16) Acetone	4.447	43	18165	32.315	ug/l	93
20) Methylene Chloride	5.294	84	11285	9.004	ug/l #	85
37) Ethyl Acetate	7.571	43	16722	8.714	ug/l #	96
43) Isopropyl Acetate	8.694	43	134549	45.201	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
Data File : VN075115.D
Acq On : 02 Nov 2022 14:27
Operator : JC\MD
Sample : N5380-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-D

Quant Time: Nov 03 04:16:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

