

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
 Data File : VN075118.D
 Acq On : 02 Nov 2022 15:39
 Operator : JC\MD
 Sample : N5396-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-B

Quant Time: Nov 03 04:18:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	101326	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	199896	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	185062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	70799	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	80364	58.864	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.720%
35) Dibromofluoromethane	8.171	113	61267	52.209	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.420%
50) Toluene-d8	10.571	98	249054	54.703	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.400%
62) 4-Bromofluorobenzene	12.853	95	77705	48.586	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.180%
Target Compounds						
						Qvalue
16) Acetone	4.453	43	10405	17.999	ug/l	95
20) Methylene Chloride	5.277	84	6714	5.209	ug/l #	81
37) Ethyl Acetate	7.571	43	15726	7.893	ug/l	97
43) Isopropyl Acetate	8.694	43	166096	53.744	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
Data File : VN075118.D
Acq On : 02 Nov 2022 15:39
Operator : JC\MD
Sample : N5396-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-B

Quant Time: Nov 03 04:18:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

