

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
 Data File : VN075124.D
 Acq On : 02 Nov 2022 18:04
 Operator : JC\MD
 Sample : N5395-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-1-103122

Quant Time: Nov 03 04:21:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	93494	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	179042	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	153371	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	57095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	76945	61.081	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	122.160%
35) Dibromofluoromethane	8.171	113	59339	56.456	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.920%
50) Toluene-d8	10.571	98	216854	53.178	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.360%
62) 4-Bromofluorobenzene	12.853	95	63304	44.192	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.380%
Target Compounds						
						Qvalue
16) Acetone	4.453	43	11510	21.579	ug/l	99
20) Methylene Chloride	5.294	84	15541	13.067	ug/l	92
95) Naphthalene	15.641	128	1262	0.388	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
Data File : VN075124.D
Acq On : 02 Nov 2022 18:04
Operator : JC\MD
Sample : N5395-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-1-103122

Quant Time: Nov 03 04:21:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

