

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
 Data File : VN075125.D
 Acq On : 02 Nov 2022 18:28
 Operator : JC\MD
 Sample : N5395-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-12-103122

Quant Time: Nov 03 04:21:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	88681	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	183731	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	153384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	60285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	79377	66.431	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	132.860%#
35) Dibromofluoromethane	8.177	113	59076	54.771	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.540%
50) Toluene-d8	10.571	98	212866	50.868	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.740%
62) 4-Bromofluorobenzene	12.853	95	65277	44.406	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.820%
Target Compounds						
						Qvalue
16) Acetone	4.453	43	12079	23.874	ug/l	97
20) Methylene Chloride	5.289	84	13828	12.258	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
Data File : VN075125.D
Acq On : 02 Nov 2022 18:28
Operator : JC\MD
Sample : N5395-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-12-103122

Quant Time: Nov 03 04:21:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

