

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
 Data File : VN075131.D
 Acq On : 02 Nov 2022 20:52
 Operator : JC\MD
 Sample : N5395-21
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-2-103122

Quant Time: Nov 03 04:24:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	91761	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	181201	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	147573	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	51127	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	77154	62.404	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 124.800%	
35) Dibromofluoromethane	8.171	113	59672	56.096	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 112.200%	
50) Toluene-d8	10.571	98	212611	51.517	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.040%	
62) 4-Bromofluorobenzene	12.853	95	58710	40.497	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 81.000%#	
Target Compounds						
16) Acetone	4.453	43	7198	13.749	ug/l	96
20) Methylene Chloride	5.295	84	10074	8.631	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
Data File : VN075131.D
Acq On : 02 Nov 2022 20:52
Operator : JC\MD
Sample : N5395-21
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-2-103122

Quant Time: Nov 03 04:24:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

