

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
 Data File : VN075132.D
 Acq On : 02 Nov 2022 21:16
 Operator : JC\MD
 Sample : N5395-22
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-6-103122

Quant Time: Nov 03 04:24:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	91519	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	174379	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	149429	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	58473	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	77079	62.508	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 125.020%#			
35) Dibromofluoromethane	8.171	113	58198	56.851	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 113.700%			
50) Toluene-d8	10.571	98	211185	53.173	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.340%			
62) 4-Bromofluorobenzene	12.853	95	62076	44.494	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 88.980%			
Target Compounds						
16) Acetone	4.453	43	2822	5.405	ug/l #	75
20) Methylene Chloride	5.283	84	14202	12.199	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
Data File : VN075132.D
Acq On : 02 Nov 2022 21:16
Operator : JC\MD
Sample : N5395-22
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-6-103122

Quant Time: Nov 03 04:24:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

