

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
 Data File : VN075133.D
 Acq On : 02 Nov 2022 21:40
 Operator : JC\MD
 Sample : N5395-23
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-7-103122

Quant Time: Nov 03 04:24:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	92923	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	175590	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	150954	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	57053	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	75523	60.321	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	120.640%
35) Dibromofluoromethane	8.177	113	57122	55.415	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.840%
50) Toluene-d8	10.571	98	212856	53.224	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.440%
62) 4-Bromofluorobenzene	12.853	95	62485	44.478	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.960%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	3828	7.221	ug/l #	77
20) Methylene Chloride	5.289	84	11515	9.742	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110222\
 Data File : VN075133.D
 Acq On : 02 Nov 2022 21:40
 Operator : JC\MD
 Sample : N5395-23
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-7-103122

Quant Time: Nov 03 04:24:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

