

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
 Data File : VN079794.D
 Acq On : 02 Nov 2023 14:00
 Operator : JC\MD
 Sample : PB156789TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156789TB

Quant Time: Nov 03 04:00:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	209278	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	418283	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	451433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	168310	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	188778	59.207	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 118.420%	
35) Dibromofluoromethane	8.165	113	153919	51.992	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.980%	
50) Toluene-d8	10.571	98	583934	53.580	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.160%	
62) 4-Bromofluorobenzene	12.853	95	192449	53.244	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.480%	
Target Compounds						Qvalue
16) Acetone	4.436	43	26245	23.434	ug/l	96
20) Methylene Chloride	5.283	84	20805	7.362	ug/l	90
43) Isopropyl Acetate	8.694	43	163078	23.987	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
Data File : VN079794.D
Acq On : 02 Nov 2023 14:00
Operator : JC\MD
Sample : PB156789TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156789TB

Quant Time: Nov 03 04:00:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

