

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
 Data File : VN079806.D
 Acq On : 02 Nov 2023 18:45
 Operator : JC\MD
 Sample : 05225-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-188-GW-340-342

Quant Time: Nov 03 04:05:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	207188	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	413324	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	429070	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	165211	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	188849	59.827	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 119.660%	
35) Dibromofluoromethane	8.171	113	157630	53.884	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.760%	
50) Toluene-d8	10.571	98	550015	51.073	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.140%	
62) 4-Bromofluorobenzene	12.853	95	178679	50.027	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.060%	
Target Compounds						
16) Acetone	4.442	43	6036	5.444	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
Data File : VN079806.D
Acq On : 02 Nov 2023 18:45
Operator : JC\MD
Sample : 05225-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-188-GW-340-342

Quant Time: Nov 03 04:05:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

