

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
 Data File : VN079807.D
 Acq On : 02 Nov 2023 19:09
 Operator : JC\MD
 Sample : 05225-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-188-GW-320-322

Quant Time: Nov 03 04:05:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	199928	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	405520	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	426591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	167410	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	191221	62.778	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 125.560%#		
35) Dibromofluoromethane	8.171	113	156268	54.447	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 108.900%		
50) Toluene-d8	10.571	98	534392	50.577	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.160%		
62) 4-Bromofluorobenzene	12.853	95	182043	51.950	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 103.900%		
Target Compounds						
16) Acetone	4.448	43	5839	5.457	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
Data File : VN079807.D
Acq On : 02 Nov 2023 19:09
Operator : JC\MD
Sample : 05225-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-188-GW-320-322

Quant Time: Nov 03 04:05:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

