

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
 Data File : VN079809.D
 Acq On : 02 Nov 2023 19:57
 Operator : JC\MD
 Sample : 05224-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-188-GW-280-282

Quant Time: Nov 03 04:06:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	205318	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	411840	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	432652	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	169274	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	190603	60.933	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 121.860%		
35) Dibromofluoromethane	8.171	113	159775	54.814	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 109.620%		
50) Toluene-d8	10.571	98	543450	50.645	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.300%		
62) 4-Bromofluorobenzene	12.853	95	184548	51.857	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 103.720%		
Target Compounds						
						Qvalue
16) Acetone	4.442	43	13435	12.227	ug/l #	85
30) Chloroform	7.977	83	4091	0.790	ug/l	86
44) Trichloroethene	9.353	130	1790	0.579	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
Data File : VN079809.D
Acq On : 02 Nov 2023 19:57
Operator : JC\MD
Sample : 05224-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-188-GW-280-282

Quant Time: Nov 03 04:06:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

