

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
 Data File : VN079820.D
 Acq On : 03 Nov 2023 03:11
 Operator : JC\MD
 Sample : 05222-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-RB-1-20231030

Quant Time: Nov 03 05:35:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	230769	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	449923	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	461734	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	176666	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	200082	56.909	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 113.820%	
35) Dibromofluoromethane	8.171	113	171161	53.750	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.500%	
50) Toluene-d8	10.570	98	593091	50.593	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.180%	
62) 4-Bromofluorobenzene	12.853	95	195312	50.236	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.480%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
Data File : VN079820.D
Acq On : 03 Nov 2023 03:11
Operator : JC\MD
Sample : 05222-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-RB-1-20231030

Quant Time: Nov 03 05:35:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

