

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN110223\  
 Data File : VN079828.D  
 Acq On : 03 Nov 2023 06:21  
 Operator : JC\MD  
 Sample : 05222-13  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 49 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 FSND-RB-3-20231101

Quant Time: Nov 03 07:46:27 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N110123W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Nov 02 05:49:08 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 8.224  | 168   | 236670   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.106  | 114   | 460826   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.871 | 117   | 470718   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.794 | 152   | 179391   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.583  | 65    | 207363   | 57.509   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =      | 115.020% |
| 35) Dibromofluoromethane    | 8.165  | 113   | 173585   | 53.222   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 106.440% |
| 50) Toluene-d8              | 10.571 | 98    | 612071   | 50.977   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =      | 101.960% |
| 62) 4-Bromofluorobenzene    | 12.853 | 95    | 196337   | 49.305   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 98.600%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 4.448  | 43    | 3891     | 3.072    | ug/l # | 82       |
| 52) Toluene                 | 10.635 | 92    | 3626     | 0.439    | ug/l   | 92       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN110223\  
Data File : VN079828.D  
Acq On : 03 Nov 2023 06:21  
Operator : JC\MD  
Sample : 05222-13  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 49 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
FSND-RB-3-20231101

Quant Time: Nov 03 07:46:27 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N110123W.M  
Quant Title : SW846 8260  
QLast Update : Thu Nov 02 05:49:08 2023  
Response via : Initial Calibration

