

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
 Data File : VN079829.D
 Acq On : 03 Nov 2023 06:44
 Operator : JC\MD
 Sample : 05222-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-16-20231101

Quant Time: Nov 03 07:46:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	212925	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	431502	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	443086	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	169466	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	195996	60.418	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	120.840%
35) Dibromofluoromethane	8.165	113	165107	54.063	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.120%
50) Toluene-d8	10.571	98	569334	50.640	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.280%
62) 4-Bromofluorobenzene	12.853	95	186113	49.914	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
Data File : VN079829.D
Acq On : 03 Nov 2023 06:44
Operator : JC\MD
Sample : 05222-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-16-20231101

Quant Time: Nov 03 07:46:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

