

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
 Data File : VN079831.D
 Acq On : 03 Nov 2023 07:32
 Operator : JC\MD
 Sample : 05222-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-EVAL-04D-20231101

Quant Time: Nov 03 08:53:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	222059	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	433489	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	452708	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	174278	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	199797	59.057	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.120%
35) Dibromofluoromethane	8.171	113	169601	55.280	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.560%
50) Toluene-d8	10.571	98	573782	50.801	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.600%
62) 4-Bromofluorobenzene	12.853	95	191713	51.180	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.360%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	4.347	96	3784	1.583	ug/l	# 65
44) Trichloroethene	9.353	130	196984	60.502	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
Data File : VN079831.D
Acq On : 03 Nov 2023 07:32
Operator : JC\MD
Sample : 05222-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-EVAL-04D-20231101

Quant Time: Nov 03 08:53:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

