

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
 Data File : VN079836.D
 Acq On : 03 Nov 2023 09:30
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 03 09:50:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.602	0.630	-4.7	96	0.00
3 P	Chloromethane	0.837	0.703	16.0	89	0.00
4 C	Vinyl Chloride	0.788	0.973	-23.5#	115	0.00
5 T	Bromomethane	0.479	0.514	-7.3	106	0.00
6 T	Chloroethane	0.623	0.594	4.7	94	0.00
7 T	Trichlorofluoromethane	0.981	0.979	0.2	95	0.00
8 T	Diethyl Ether	0.372	0.378	-1.6	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.568	0.561	1.2	96	0.00
10 T	Methyl Iodide	0.649	0.652	-0.5	89	0.00
11 T	Tert butyl alcohol	0.097	0.090	7.2	91	0.00
12 CM	1,1-Dichloroethene	0.538	0.528	1.9#	91	0.00
13 T	Acrolein	0.080	0.075	6.3	91	0.00
14 T	Allyl chloride	0.824	0.834	-1.2	97	0.00
15 T	Acrylonitrile	0.303	0.307	-1.3	97	0.00
16 T	Acetone	0.268	0.238	11.2	90	0.00
17 T	Carbon Disulfide	1.552	1.507	2.9	94	0.00
18 T	Methyl Acetate	1.280	1.298	-1.4	101	0.00
19 T	Methyl tert-butyl Ether	1.853	1.912	-3.2	95	0.00
20 T	Methylene Chloride	0.675	0.667	1.2	98	0.01
21 T	trans-1,2-Dichloroethene	0.600	0.611	-1.8	98	0.00
22 T	Diisopropyl ether	1.910	2.168	-13.5	103	0.00
23 T	Vinyl Acetate	1.054	1.119	-6.2	94	0.00
24 P	1,1-Dichloroethane	1.161	1.241	-6.9	102	0.00
25 T	2-Butanone	0.392	0.405	-3.3	95	0.00
26 T	2,2-Dichloropropane	0.960	0.776	19.2	75	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.744	-3.6	99	0.00
28 T	Bromochloromethane	0.566	0.639	-12.9	106	0.00
29 T	Tetrahydrofuran	0.255	0.272	-6.7	99	0.00
30 C	Chloroform	1.261	1.265	-0.3#	101	0.00
31 T	Cyclohexane	1.007	1.021	-1.4	101	0.00
32 T	1,1,1-Trichloroethane	1.028	1.088	-5.8	101	0.00
33 S	1,2-Dichloroethane-d4	0.762	0.792	-3.9	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.354	0.358	-1.1	96	0.00
36 T	1,1-Dichloropropene	0.505	0.538	-6.5	103	0.00
37 T	Ethyl Acetate	0.463	0.462	0.2	94	0.00
38 T	Carbon Tetrachloride	0.509	0.533	-4.7	101	0.00
39 T	Methylcyclohexane	0.476	0.507	-6.5	97	0.00
40 TM	Benzene	1.513	1.571	-3.8	102	0.00
41 T	Methacrylonitrile	0.246	0.272	-10.6	105	0.00
42 TM	1,2-Dichloroethane	0.529	0.578	-9.3	107	0.00
43 T	Isopropyl Acetate	0.813	0.806	0.9	96	0.00
44 TM	Trichloroethene	0.376	0.379	-0.8	99	0.00
45 C	1,2-Dichloropropane	0.387	0.425	-9.8#	105	0.00
46 T	Dibromomethane	0.264	0.276	-4.5	102	0.00
47 T	Bromodichloromethane	0.563	0.571	-1.4	101	0.00
48 T	Methyl methacrylate	0.356	0.400	-12.4	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
 Data File : VN079836.D
 Acq On : 03 Nov 2023 09:30
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 03 09:50:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	89	0.00
50 S	Toluene-d8	1.303	1.313	-0.8	93	0.00
51 T	4-Methyl-2-Pentanone	0.458	0.494	-7.9	100	0.00
52 CM	Toluene	0.897	0.971	-8.2#	100	0.00
53 T	t-1,3-Dichloropropene	0.552	0.570	-3.3	94	0.00
54 T	cis-1,3-Dichloropropene	0.602	0.629	-4.5	94	0.00
55 T	1,1,2-Trichloroethane	0.361	0.382	-5.8	101	0.00
56 T	Ethyl methacrylate	0.515	0.559	-8.5	95	0.00
57 T	1,3-Dichloropropane	0.636	0.682	-7.2	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.261	0.271	-3.8	92	0.00
59 T	2-Hexanone	0.350	0.351	-0.3	93	0.00
60 T	Dibromochloromethane	0.386	0.401	-3.9	96	0.00
61 T	1,2-Dibromoethane	0.362	0.368	-1.7	95	0.00
62 S	4-Bromofluorobenzene	0.432	0.443	-2.5	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.324	0.378	-16.7	111	0.00
65 PM	Chlorobenzene	1.085	1.110	-2.3	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.391	0.424	-8.4	102	0.00
67 C	Ethyl Benzene	1.890	2.064	-9.2#	99	0.00
68 T	m/p-Xylenes	0.704	0.774	-9.9	100	0.00
69 T	o-Xylene	0.678	0.743	-9.6	99	0.00
70 T	Styrene	1.089	1.259	-15.6	99	0.00
71 P	Bromoform	0.277	0.302	-9.0	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.795	3.946	-4.0	99	0.00
74 T	N-amyl acetate	1.656	1.600	3.4	91	0.00
75 P	1,1,2,2-Tetrachloroethane	1.345	1.261	6.2	97	0.00
76 T	1,2,3-Trichloropropane	1.174	1.100	6.3	96	0.00
77 T	Bromobenzene	0.929	0.931	-0.2	99	0.00
78 T	n-propylbenzene	4.346	4.694	-8.0	99	0.00
79 T	2-Chlorotoluene	2.776	2.832	-2.0	100	0.00
80 T	1,3,5-Trimethylbenzene	2.952	3.259	-10.4	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.431	0.348	19.3	72	0.00
82 T	4-Chlorotoluene	2.759	2.841	-3.0	99	0.00
83 T	tert-Butylbenzene	2.417	2.657	-9.9	100	0.00
84 T	1,2,4-Trimethylbenzene	2.905	3.239	-11.5	100	0.00
85 T	sec-Butylbenzene	3.348	3.706	-10.7	101	0.00
86 T	p-Isopropyltoluene	2.730	3.085	-13.0	101	0.00
87 T	1,3-Dichlorobenzene	1.728	1.733	-0.3	99	0.00
88 T	1,4-Dichlorobenzene	1.754	1.675	4.5	96	0.00
89 T	n-Butylbenzene	2.351	2.587	-10.0	98	0.00
90 T	Hexachloroethane	0.541	0.579	-7.0	110	0.00
91 T	1,2-Dichlorobenzene	1.632	1.638	-0.4	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.228	0.224	1.8	92	0.00
93 T	1,2,4-Trichlorobenzene	0.776	0.805	-3.7	96	0.00
94 T	Hexachlorobutadiene	0.352	0.364	-3.4	105	0.00
95 T	Naphthalene	2.631	2.434	7.5	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110223\
Data File : VN079836.D
Acq On : 03 Nov 2023 09:30
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 03 09:50:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.760	0.780	-2.6	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6