

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
 Data File : VN075137.D
 Acq On : 03 Nov 2022 12:13
 Operator : JC\MD
 Sample : VN1103MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1103MBL01

Quant Time: Nov 04 04:09:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	95229	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	177182	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	149896	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	57387	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	74527	58.084	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.160%
35) Dibromofluoromethane	8.171	113	55437	53.297	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.600%
50) Toluene-d8	10.571	98	208109	51.570	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.140%
62) 4-Bromofluorobenzene	12.853	95	62777	44.284	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
Data File : VN075137.D
Acq On : 03 Nov 2022 12:13
Operator : JC\MD
Sample : VN1103MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1103MBL01

Quant Time: Nov 04 04:09:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

