

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
 Data File : VN075141.D
 Acq On : 03 Nov 2022 14:00
 Operator : JC\MD
 Sample : N5395-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-12-103122

Quant Time: Nov 04 04:10:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	94871	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	183442	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	152850	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	60147	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	75124	58.770	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 117.540%			
35) Dibromofluoromethane	8.171	113	58213	54.056	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.120%			
50) Toluene-d8	10.571	98	216950	51.926	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.860%			
62) 4-Bromofluorobenzene	12.853	95	63974	43.589	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 87.180%			
Target Compounds						Qvalue
16) Acetone	4.442	43	4025	7.436	ug/l #	81
20) Methylene Chloride	5.289	84	10666	8.838	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
Data File : VN075141.D
Acq On : 03 Nov 2022 14:00
Operator : JC\MD
Sample : N5395-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-12-103122

Quant Time: Nov 04 04:10:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

