

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
 Data File : VN075142.D
 Acq On : 03 Nov 2022 14:24
 Operator : JC\MD
 Sample : N5395-17RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-10-103122RE

Quant Time: Nov 04 04:10:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	91509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	181334	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	159420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	61931	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	80081	64.949	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 129.900%#	
35) Dibromofluoromethane	8.171	113	59135	55.551	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.100%	
50) Toluene-d8	10.571	98	220954	53.499	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.000%	
62) 4-Bromofluorobenzene	12.853	95	65848	45.387	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.780%	
Target Compounds						
						Qvalue
16) Acetone	4.459	43	1733	3.319	ug/l	95
20) Methylene Chloride	5.277	84	15814	13.585	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
Data File : VN075142.D
Acq On : 03 Nov 2022 14:24
Operator : JC\MD
Sample : N5395-17RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-10-103122RE

Quant Time: Nov 04 04:10:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

