

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
 Data File : VN075143.D
 Acq On : 03 Nov 2022 14:49
 Operator : JC\MD
 Sample : N5395-19RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-5-103122RE

Quant Time: Nov 04 04:10:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	94091	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	186828	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	159354	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	62953	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	80293	63.334	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 126.660%#	
35) Dibromofluoromethane	8.171	113	61634	56.196	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 112.400%	
50) Toluene-d8	10.571	98	225810	53.067	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.140%	
62) 4-Bromofluorobenzene	12.853	95	68867	46.072	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.140%	
Target Compounds						Qvalue
16) Acetone	4.448	43	5104	9.508	ug/l	89
20) Methylene Chloride	5.283	84	13588	11.353	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
Data File : VN075143.D
Acq On : 03 Nov 2022 14:49
Operator : JC\MD
Sample : N5395-19RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-5-103122RE

Quant Time: Nov 04 04:10:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

