

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
 Data File : VN075146.D
 Acq On : 03 Nov 2022 16:01
 Operator : JC\MD
 Sample : N5395-25
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-9-103122

Quant Time: Nov 04 04:10:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	89593	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	173659	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	141168	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	55277	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	72495	60.054	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	120.100%
35) Dibromofluoromethane	8.171	113	55770	54.705	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.420%
50) Toluene-d8	10.571	98	199654	50.478	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.960%
62) 4-Bromofluorobenzene	12.853	95	59932	43.135	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.260%
Target Compounds						
						Qvalue
16) Acetone	4.447	43	2903	5.679	ug/l	# 88
20) Methylene Chloride	5.289	84	8688	7.623	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
Data File : VN075146.D
Acq On : 03 Nov 2022 16:01
Operator : JC\MD
Sample : N5395-25
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-9-103122

Quant Time: Nov 04 04:10:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

