

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
 Data File : VN075148.D
 Acq On : 03 Nov 2022 16:49
 Operator : JC\MD
 Sample : PB148771TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148771TB

Quant Time: Nov 04 04:11:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	91634	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	179975	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	169427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	64047	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	73524	59.550	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.100%
35) Dibromofluoromethane	8.177	113	56552	53.525	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.060%
50) Toluene-d8	10.571	98	222804	54.354	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.700%
62) 4-Bromofluorobenzene	12.853	95	70176	48.735	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.480%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	10376	19.848	ug/l	95
20) Methylene Chloride	5.277	84	7990	6.855	ug/l #	78
37) Ethyl Acetate	7.571	43	14806	8.254	ug/l	97
43) Isopropyl Acetate	8.694	43	66092	23.753	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
Data File : VN075148.D
Acq On : 03 Nov 2022 16:49
Operator : JC\MD
Sample : PB148771TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148771TB

Quant Time: Nov 04 04:11:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

