

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
 Data File : VN075150.D
 Acq On : 03 Nov 2022 17:37
 Operator : JC\MD
 Sample : PB148771ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148771ZHE#02

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/04/2022
 Supervised By : Mahesh Dadoda 11/04/2022

Quant Time: Nov 04 04:11:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	90000	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	186762	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	168675	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	63374	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	75879	62.573	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 125.140%#			
35) Dibromofluoromethane	8.171	113	58787	53.619	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.240%			
50) Toluene-d8	10.571	98	227082	53.385	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.760%			
62) 4-Bromofluorobenzene	12.847	95	71856	48.089	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.180%			
Target Compounds						Qvalue
16) Acetone	4.448	43	8923	17.378	ug/l	99
20) Methylene Chloride	5.277	84	8478	7.405	ug/l	88
37) Ethyl Acetate	7.571	43	15938	8.562	ug/l	96
43) Isopropyl Acetate	8.694	43	146704m	50.808	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
Data File : VN075150.D
Acq On : 03 Nov 2022 17:37
Operator : JC\MD
Sample : PB148771ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148771ZHE#02

Manual Integrations
APPROVED

Reviewed By : John Carlone 11/04/2022
Supervised By : Mahesh Dadoda 11/04/2022

Quant Time: Nov 04 04:11:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

