

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
 Data File : VN075151.D
 Acq On : 03 Nov 2022 18:01
 Operator : JC\MD
 Sample : PB148771ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148771ZHE#03

Quant Time: Nov 04 04:11:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	89745	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	184239	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	168083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	61628	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	75488	62.428	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 124.860%		
35) Dibromofluoromethane	8.171	113	58495	54.083	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 108.160%		
50) Toluene-d8	10.571	98	227677	54.258	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 108.520%		
62) 4-Bromofluorobenzene	12.853	95	71337	48.395	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 96.800%		
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	9581	18.713	ug/l	99
20) Methylene Chloride	5.289	84	8682	7.605	ug/l #	90
37) Ethyl Acetate	7.571	43	16334	8.895	ug/l	97
43) Isopropyl Acetate	8.694	43	179308	62.950	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
Data File : VN075151.D
Acq On : 03 Nov 2022 18:01
Operator : JC\MD
Sample : PB148771ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148771ZHE#03

Quant Time: Nov 04 04:11:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

