

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
 Data File : VN075153.D
 Acq On : 03 Nov 2022 18:49
 Operator : JC\MD
 Sample : PB148771ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148771ZHE#05

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/04/2022
 Supervised By : Mahesh Dadoda 11/04/2022

Quant Time: Nov 04 04:11:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	92143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	186305	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	173284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	66110	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	76524	61.637	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 123.280%			
35) Dibromofluoromethane	8.171	113	58503	53.491	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.980%			
50) Toluene-d8	10.571	98	233621	55.057	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 110.120%			
62) 4-Bromofluorobenzene	12.853	95	75010	50.323	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 100.640%			
Target Compounds						
16) Acetone	4.453	43	8965	17.054	ug/l	89
18) Methyl Acetate	5.041	43	4113	2.395	ug/l	97
20) Methylene Chloride	5.289	84	8221	7.014	ug/l #	90
37) Ethyl Acetate	7.571	43	15477	8.335	ug/l	98
43) Isopropyl Acetate	8.688	43	172200m	59.784	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
Data File : VN075153.D
Acq On : 03 Nov 2022 18:49
Operator : JC\MD
Sample : PB148771ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148771ZHE#05

Manual Integrations
APPROVED

Reviewed By : John Carlone 11/04/2022
Supervised By : Mahesh Dadoda 11/04/2022

Quant Time: Nov 04 04:11:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

