

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
 Data File : VN075154.D
 Acq On : 03 Nov 2022 19:13
 Operator : JC\MD
 Sample : PB148771ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148771ZHE#06

Quant Time: Nov 04 04:12:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	94345	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	189691	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	177014	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	67473	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	76656	60.303	ug/l	0.00
Spiked Amount 50.000	Range	74 - 125	Recovery	= 120.600%		
35) Dibromofluoromethane	8.171	113	59338	53.286	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	= 106.580%		
50) Toluene-d8	10.571	98	235589	54.529	ug/l	0.00
Spiked Amount 50.000	Range	86 - 113	Recovery	= 109.060%		
62) 4-Bromofluorobenzene	12.853	95	74567	49.132	ug/l	0.00
Spiked Amount 50.000	Range	83 - 123	Recovery	= 98.260%		
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	8994	16.710	ug/l	90
18) Methyl Acetate	5.047	43	4228	2.405	ug/l #	81
20) Methylene Chloride	5.294	84	9072	7.559	ug/l #	76
37) Ethyl Acetate	7.565	43	15501	8.199	ug/l	99
43) Isopropyl Acetate	8.688	43	191897	65.433	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
Data File : VN075154.D
Acq On : 03 Nov 2022 19:13
Operator : JC\MD
Sample : PB148771ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148771ZHE#06

Quant Time: Nov 04 04:12:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

