

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
 Data File : VN075156.D
 Acq On : 03 Nov 2022 20:01
 Operator : JC\MD
 Sample : N5418-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-J

Quant Time: Nov 04 04:12:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/04/2022
 Supervised By : Mahesh Dadoda 11/04/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	93150	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	187340	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	172900	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	65944	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	75688	60.305	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 120.620%			
35) Dibromofluoromethane	8.171	113	57873	52.622	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.240%			
50) Toluene-d8	10.571	98	232849	54.572	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 109.140%			
62) 4-Bromofluorobenzene	12.853	95	72413	48.312	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.620%			
Target Compounds					Qvalue	
16) Acetone	4.453	43	12719	23.933	ug/l	97
18) Methyl Acetate	5.047	43	4861	2.800	ug/l #	88
20) Methylene Chloride	5.283	84	11213	9.463	ug/l #	86
37) Ethyl Acetate	7.571	43	14616	7.827	ug/l	98
43) Isopropyl Acetate	8.694	43	146449m	50.563	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
Data File : VN075156.D
Acq On : 03 Nov 2022 20:01
Operator : JC\MD
Sample : N5418-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-J

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/04/2022
Supervised By :Mahesh Dadoda 11/04/2022

Quant Time: Nov 04 04:12:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

