

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
 Data File : VN075158.D
 Acq On : 03 Nov 2022 20:49
 Operator : JC\MD
 Sample : N5418-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-L

Quant Time: Nov 04 04:12:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/04/2022
 Supervised By : Mahesh Dadoda 11/04/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	91958	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	190239	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	171918	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	64602	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	76599	61.822	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 123.640%			
35) Dibromofluoromethane	8.171	113	59595	53.362	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.720%			
50) Toluene-d8	10.571	98	233770	53.953	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.900%			
62) 4-Bromofluorobenzene	12.853	95	71695	47.104	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.200%			
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	11100	21.158	ug/l #	88
18) Methyl Acetate	5.047	43	4068	2.374	ug/l	99
20) Methylene Chloride	5.289	84	10925	9.340	ug/l	91
37) Ethyl Acetate	7.571	43	16345	8.620	ug/l	96
43) Isopropyl Acetate	8.694	43	148216m	50.393	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110322\
Data File : VN075158.D
Acq On : 03 Nov 2022 20:49
Operator : JC\MD
Sample : N5418-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-L

Quant Time: Nov 04 04:12:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/04/2022
Supervised By :Mahesh Dadoda 11/04/2022

