

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110323\
 Data File : VN079843.D
 Acq On : 03 Nov 2023 12:53
 Operator : JC\MD
 Sample : 05224-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-188-GW-260-262

Quant Time: Nov 04 00:51:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	247152	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	487962	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	479164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	185522	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	200993	53.378	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.760%
35) Dibromofluoromethane	8.171	113	165827	48.016	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.040%
50) Toluene-d8	10.571	98	592409	46.595	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.200%
62) 4-Bromofluorobenzene	12.853	95	196728	46.656	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.320%
Target Compounds						
						Qvalue
44) Trichloroethene	9.353	130	19570	5.340	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110323\
Data File : VN079843.D
Acq On : 03 Nov 2023 12:53
Operator : JC\MD
Sample : 05224-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-188-GW-260-262

Quant Time: Nov 04 00:51:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

