

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110323\
 Data File : VN079881.D
 Acq On : 04 Nov 2023 06:43
 Operator : JC\MD
 Sample : 05223-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-12R-20231031

Quant Time: Nov 06 00:28:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	234194	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	469961	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	493765	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	195172	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	210209	58.915	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 117.820%			
35) Dibromofluoromethane	8.171	113	171860	51.669	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.340%			
50) Toluene-d8	10.571	98	595714	48.650	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.300%			
62) 4-Bromofluorobenzene	12.853	95	199400	49.101	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 98.200%			
Target Compounds						Qvalue
38) Carbon Tetrachloride	8.365	117	15649	3.269	ug/l	99
44) Trichloroethene	9.353	130	23052	6.531	ug/l	93
64) Tetrachloroethene	11.106	164	2092	0.654	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110323\
Data File : VN079881.D
Acq On : 04 Nov 2023 06:43
Operator : JC\MD
Sample : 05223-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-12R-20231031

Quant Time: Nov 06 00:28:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

