

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110323\
 Data File : VN079882.D
 Acq On : 04 Nov 2023 07:07
 Operator : JC\MD
 Sample : 05223-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-20-20231030

Quant Time: Nov 06 00:28:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	242052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	479993	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	493720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	185924	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	206001	55.861	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.720%
35) Dibromofluoromethane	8.165	113	167994	49.451	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.900%
50) Toluene-d8	10.571	98	596608	47.705	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.400%
62) 4-Bromofluorobenzene	12.853	95	194544	46.904	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.800%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.353	130	235117	65.218	ug/l	97
64) Tetrachloroethene	11.106	164	8081	2.527	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110323\
Data File : VN079882.D
Acq On : 04 Nov 2023 07:07
Operator : JC\MD
Sample : 05223-13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-20-20231030

Quant Time: Nov 06 00:28:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

