

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110323\
 Data File : VN079883.D
 Acq On : 04 Nov 2023 07:31
 Operator : JC\MD
 Sample : 05223-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-BR-1-20231030

Quant Time: Nov 06 00:28:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	248372	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	499173	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	508679	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	197571	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	217920	57.589	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.180%
35) Dibromofluoromethane	8.171	113	182475	51.650	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.300%
50) Toluene-d8	10.571	98	616005	47.363	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.720%
62) 4-Bromofluorobenzene	12.853	95	207044	48.000	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110323\
Data File : VN079883.D
Acq On : 04 Nov 2023 07:31
Operator : JC\MD
Sample : 05223-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-BR-1-20231030

Quant Time: Nov 06 00:28:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

