

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN110420\
 Data File : VN064436.D
 Acq On : 4 Nov 2020 15:12
 Operator : JC/MD
 Sample : L4215-01
 Misc : 5.41g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AU-05-102820

Quant Time: Nov 05 01:24:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	233965	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	459072	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	451824	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	183361	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	210712	51.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.40%	
35) Dibromofluoromethane	7.55	113	153455	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
50) Toluene-d8	10.06	98	580807	48.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.16%	
62) 4-Bromofluorobenzene	12.38	95	227264	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110420\
Data File : VN064436.D
Acq On : 4 Nov 2020 15:12
Operator : JC/MD
Sample : L4215-01
Misc : 5.41g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AU-05-102820

Quant Time: Nov 05 01:24:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

