

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110420\
 Data File : VN064439.D
 Acq On : 4 Nov 2020 16:24
 Operator : JC/MD
 Sample : L4627-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB178-TB-20201102

Quant Time: Nov 05 01:32:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	269563	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	527166	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	538559	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	202835	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	245142	52.21	ug/l	0.00
Spiked Amount			Recovery	=	104.42%	
35) Dibromofluoromethane	7.55	113	171676	47.83	ug/l	0.00
Spiked Amount			Recovery	=	95.66%	
50) Toluene-d8	10.06	98	681296	49.11	ug/l	0.00
Spiked Amount			Recovery	=	98.22%	
62) 4-Bromofluorobenzene	12.38	95	265481	50.60	ug/l	0.00
Spiked Amount			Recovery	=	101.20%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110420\
Data File : VN064439.D
Acq On : 4 Nov 2020 16:24
Operator : JC/MD
Sample : L4627-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB178-TB-20201102

Quant Time: Nov 05 01:32:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

