

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110420\
 Data File : VN064445.D
 Acq On : 4 Nov 2020 18:50
 Operator : JC/MD
 Sample : PB132810TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132810TB

Quant Time: Nov 05 01:46:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	251024	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	496934	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	538933	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	202981	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	239895	54.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.74%	
35) Dibromofluoromethane	7.55	113	168705	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
50) Toluene-d8	10.06	98	660355	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
62) 4-Bromofluorobenzene	12.38	95	263871	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
Target Compounds						
16) Acetone	3.78	43	17194	10.323	ug/l	99
18) Methyl Acetate	4.29	43	4534	1.385	ug/l #	74
20) Methylene Chloride	4.51	84	13373	4.047	ug/l #	91
43) Isopropyl Acetate	8.13	43	107278	12.610	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110420\
Data File : VN064445.D
Acq On : 4 Nov 2020 18:50
Operator : JC/MD
Sample : PB132810TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132810TB

Quant Time: Nov 05 01:46:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

