

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110420\
 Data File : VN064446.D
 Acq On : 4 Nov 2020 19:14
 Operator : JC/MD
 Sample : PB132810ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132810ZHE#01

Quant Time: Nov 05 01:48:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	244710	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	488357	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	516103	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	195194	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	229652	53.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.76%	
35) Dibromofluoromethane	7.55	113	163222	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
50) Toluene-d8	10.06	98	646750	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
62) 4-Bromofluorobenzene	12.38	95	253287	52.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.22%	
Target Compounds						
16) Acetone	3.78	43	17214	10.602	ug/l	99
18) Methyl Acetate	4.29	43	3945	1.236	ug/l #	65
20) Methylene Chloride	4.50	84	13107	4.069	ug/l	92
43) Isopropyl Acetate	8.13	43	106077	12.687	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110420\
Data File : VN064446.D
Acq On : 4 Nov 2020 19:14
Operator : JC/MD
Sample : PB132810ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132810ZHE#01

Quant Time: Nov 05 01:48:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

