

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110422\
 Data File : VN075166.D
 Acq On : 04 Nov 2022 11:29
 Operator : JC\MD
 Sample : VN1104WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1104WBL01

Quant Time: Nov 05 06:55:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	95705	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	180287	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	154107	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	59993	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	76239	59.122	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.240%
35) Dibromofluoromethane	8.171	113	59158	55.895	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.800%
50) Toluene-d8	10.571	98	214550	52.250	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.500%
62) 4-Bromofluorobenzene	12.847	95	63956	44.339	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110422\
Data File : VN075166.D
Acq On : 04 Nov 2022 11:29
Operator : JC\MD
Sample : VN1104WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1104WBL01

Quant Time: Nov 05 06:55:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

