

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110422\
 Data File : VN075170.D
 Acq On : 04 Nov 2022 13:21
 Operator : JC\MD
 Sample : N5418-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-N

Quant Time: Nov 05 06:58:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	98737	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	192416	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	181105	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	69303	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	78275	58.837	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.680%
35) Dibromofluoromethane	8.171	113	60493	53.554	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.100%
50) Toluene-d8	10.571	98	241630	55.136	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.280%
62) 4-Bromofluorobenzene	12.853	95	76593	49.753	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.500%
Target Compounds						
						Qvalue
16) Acetone	4.447	43	10556	18.739	ug/l	94
18) Methyl Acetate	5.047	43	2439	1.326	ug/l #	53
20) Methylene Chloride	5.283	84	11534	9.183	ug/l	94
37) Ethyl Acetate	7.571	43	16769	8.744	ug/l	99
43) Isopropyl Acetate	8.694	43	102512	34.459	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110422\
Data File : VN075170.D
Acq On : 04 Nov 2022 13:21
Operator : JC\MD
Sample : N5418-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-N

Quant Time: Nov 05 06:58:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

