

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110422\
 Data File : VN075172.D
 Acq On : 04 Nov 2022 14:09
 Operator : JC\MD
 Sample : PB148795ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148795ZHE#01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/07/2022
 Supervised By : Mahesh Dadoda 11/07/2022

Quant Time: Nov 05 06:58:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	97198	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	200868	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	182567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	72095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	77654	59.295	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 118.580%			
35) Dibromofluoromethane	8.171	113	60701	51.477	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.960%			
50) Toluene-d8	10.571	98	243328	53.187	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.380%			
62) 4-Bromofluorobenzene	12.847	95	79764	49.632	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 99.260%			
Target Compounds					Qvalue	
16) Acetone	4.459	43	8423	15.189	ug/l	96
18) Methyl Acetate	5.042	43	4389	2.423	ug/l	93
20) Methylene Chloride	5.277	84	9252	7.483	ug/l	94
37) Ethyl Acetate	7.571	43	17282	8.632	ug/l	98
43) Isopropyl Acetate	8.694	43	142093m	45.755	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110422\
Data File : VN075172.D
Acq On : 04 Nov 2022 14:09
Operator : JC\MD
Sample : PB148795ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148795ZHE#01

Quant Time: Nov 05 06:58:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/07/2022
Supervised By :Mahesh Dadoda 11/07/2022

