

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110422\
 Data File : VN075180.D
 Acq On : 04 Nov 2022 17:41
 Operator : JC\MD
 Sample : N5482-02 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FLOOR-DRAIN-WATER

Quant Time: Nov 05 07:01:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	78586	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	149030	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	125789	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	50503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	62020	58.573	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 117.140%	
35) Dibromofluoromethane	8.177	113	46728	53.411	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.820%	
50) Toluene-d8	10.571	98	170888	50.346	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.700%	
62) 4-Bromofluorobenzene	12.853	95	52561	44.082	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.160%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110422\
Data File : VN075180.D
Acq On : 04 Nov 2022 17:41
Operator : JC\MD
Sample : N5482-02 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FLOOR-DRAIN-WATER

Quant Time: Nov 05 07:01:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

