

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110423\
 Data File : VN079890.D
 Acq On : 04 Nov 2023 14:06
 Operator : JC\MD
 Sample : VN1104MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1104MBL01

Quant Time: Nov 06 02:12:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	254501	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	498903	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	508387	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	201516	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	212506	54.806	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.620%
35) Dibromofluoromethane	8.165	113	174071	49.297	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.600%
50) Toluene-d8	10.571	98	612966	47.155	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.300%
62) 4-Bromofluorobenzene	12.853	95	209471	48.588	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110423\
Data File : VN079890.D
Acq On : 04 Nov 2023 14:06
Operator : JC\MD
Sample : VN1104MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1104MBL01

Quant Time: Nov 06 02:12:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

