

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110423\
 Data File : VN079894.D
 Acq On : 04 Nov 2023 15:41
 Operator : JC\MD
 Sample : 05222-05DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-29-20231101DL

Quant Time: Nov 06 02:14:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	239540	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	478176	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	493340	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	191665	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	211276	57.892	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 115.780%			
35) Dibromofluoromethane	8.171	113	172279	50.905	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.800%			
50) Toluene-d8	10.571	98	593125	47.606	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.220%			
62) 4-Bromofluorobenzene	12.853	95	203787	49.319	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 98.640%			
Target Compounds						Qvalue
44) Trichloroethene	9.353	130	147820	41.159	ug/l	99
64) Tetrachloroethene	11.106	164	15073	4.717	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110423\
Data File : VN079894.D
Acq On : 04 Nov 2023 15:41
Operator : JC\MD
Sample : 05222-05DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-29-20231101DL

Quant Time: Nov 06 02:14:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

