

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110423\
 Data File : VN079898.D
 Acq On : 04 Nov 2023 17:17
 Operator : JC\MD
 Sample : 05222-15DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-EVAL-04S-20231101DL

Quant Time: Nov 06 02:16:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	244555	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	493317	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	506810	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	184812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	218899	58.751	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.500%
35) Dibromofluoromethane	8.171	113	178607	51.155	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.300%
50) Toluene-d8	10.571	98	623234	48.488	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.980%
62) 4-Bromofluorobenzene	12.853	95	200041	46.927	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.860%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.353	130	81453	21.983	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110423\
Data File : VN079898.D
Acq On : 04 Nov 2023 17:17
Operator : JC\MD
Sample : 05222-15DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-EVAL-04S-20231101DL

Quant Time: Nov 06 02:16:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

