

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110423\
 Data File : VN079902.D
 Acq On : 04 Nov 2023 18:52
 Operator : JC\MD
 Sample : 05223-02DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-GEP-3-20231031DL

Quant Time: Nov 06 02:18:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	231520	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	467919	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	489061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	192661	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	211029	59.828	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.660%
35) Dibromofluoromethane	8.171	113	175216	52.908	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.820%
50) Toluene-d8	10.571	98	598690	49.106	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.220%
62) 4-Bromofluorobenzene	12.853	95	202636	50.115	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.240%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.359	130	116265	33.082	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110423\
Data File : VN079902.D
Acq On : 04 Nov 2023 18:52
Operator : JC\MD
Sample : 05223-02DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-GEP-3-20231031DL

Quant Time: Nov 06 02:18:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

